


# Method Development and Research for Measuring Emissions from Spray Polyurethane Foam (SPF) Insulation

---

**John Sebroski, Bayer MaterialScience LLC**  
Materials Characterization, Environmental Analytics

---



Center for the  
Polyurethanes Industry

# Acknowledgements

---

**Environmental Analytics-** Tom Frampton, Karen Mattson, Jason Miller, Luann Bidka-Rieck, John Sebroski, Carl Thompson

**Machine Shop / Fabrication-** Michael Bellaire

**Spray Foam Applicator-** William Pate

**Technical Operations-** Ronald Zibert

**Product Safety-** Scott Ecoff, Brian Karlovich, Jim Chapman

**Product Development-** Jim Lambach

**Vendor / Analytical Support-** Paul Morris and Caroline Widdowson, Markes International

**Materials for Generic Formulations-** Todd Wishneski, BASF

# Method Development and Research

---

Generic Foam Formulations and Compounds of Interest

Evaluation of Thermal Desorption GC/MS for Measuring Potential SPF Emissions

Micro Chamber Parameters and Specimen Preparation

SPF Sample Holding Time and Packaging Evaluation

Evaluation of Wall Effects of Semi-Volatile Compounds in Test Chambers (Adhesion to Chamber Walls)

# Generic SPF Formulations

---

Three generic SPF formulations were developed to evaluate the test methods for measuring emissions.

The formulations represent the following sample types:

- Open-cell, low density (1/2 pound) high pressure SPF
- Closed-cell, medium density (2 pound) high pressure SPF
- Kit formulation SPF, 2 components, low-pressure

# Target Compound List for Generic SPF Formulations

---

Target Compound	Acronym	Description
HFC-245a	-	Blowing Agent
HFC-134a	-	Blowing Agent
Tris-(1-chloro-2-propyl) phosphate	TCPP	Flame Retardant
Bis (2-Dimethylaminoethyl) ether	BDMAEE	Catalyst
Tetramethyliminobispropylamine	TMIBPA	Catalyst
N,N,N-Trimethylaminoethylethanolamine	TMAEEA	Catalyst
Pentamethyldiethylene triamine	PMDTA	Catalyst
Bis (dimethylaminopropyl) methylamine	DAPA	Catalyst
Methylene diphenyl diisocyanate	MDI	Isocyanate

# Analytical Method Development

## Thermal Desorption, GC/MS

---





<b>TD:</b>	<b>Markes TD-100</b>
<b>Flow path temperature</b>	160 °C
<b>Sorbent Tube</b>	Tenax TA and Carboxpack X sorbent tubes (Stainless Steel)
<b>Split in Standby</b>	10 ml/min
<b>Cold Trap</b>	U-T12ME-2S, Materials Emissions Trap with Tenax and Carbograph 5 TD sorbent
<b>Dry Purge</b>	1 min, 20 ml/min flow to split
<b>Prepurge</b>	0.1 min, default
<b>Primary desorption</b>	270 °C for 8 min, 35 ml/min trap flow, no split flow
<b>Pre-Trap Fire Purge</b>	1 min, 35 ml/min trap flow, 50 ml/min split flow
<b>Cold trap conditions</b>	Trap low: 25 °C, trap high: 300 °C, heating rate: MAX, hold time: 3 min, 50 ml/min split flow
<b>Overall TD split</b>	34.3:1

# TD-GC/MS Instrument Conditions



<b>GC/MS:</b>	<b>Agilent 7890 GC and 5975 MSD</b>
<b>Column</b>	Restek Rtx-5 Amine, 30 m, 0.25 mm x 0.5 $\mu$ m
<b>Column flow</b>	1.5 ml/min, constant flow
<b>Temperature program</b>	40 °C (2 min), 20 °C/min to 300 °C (2 min)
<b>Total run time</b>	17.0 min
<b>Carrier gas</b>	He
<b>GC inlet temperature</b>	200 °C
<b>MS source temperature</b>	230 °C
<b>MS quad temperature</b>	150 °C
<b>MS transfer line temperature</b>	250 °C
<b>Mass scan range</b>	m/z = 50-550amu

# Example- Total Ion Chromatograms





# Target Compounds by TD-GC/MS

---

Peak Number	Compound Name	MS Quantitation Ion	Retention Time, minutes	Estimated Quantitation Limit, ng
1	HFC-245a	51	1.485	20
2	TMAEEA	58	7.691	200
3	BDMAEE	58	7.727	20
4	PMDTA	72	8.265	50
5	DAPA	58	9.560	100
6	TMIBPA	58	9.654	200
7	TCPP	99	12.477	50

# TD-GC/MS Precision and Recovery

---

Compound	True Value	Mean	Recovery, %	Standard Deviation	RSD, %
HFC-245a	706	634	89.7	69.1	10.9
TMAEEA	2179	1559	71.5	383	24.6
BDMAEE	1034	977	94.5	40.1	4.10
PMDTA	1114	1035	92.9	36.4	3.52
DAPA	1072	1031	96.2	34.8	3.38
TMIBPA	2397	1631	68.1	217	13.3
TCPP	996	922	92.6	27.1	2.94

Unless otherwise stated, results reported in nanograms (ng) per thermal desorption (TD) tube

# Micro Chamber- Specimen Preparation


---



Closed-cell SPF specimen fits tightly into the micro chamber.

# Open-Cell and Kit Formulations

---



Open-cell SPF with aluminum shim ring fits tightly into the micro chamber.



Low pressure kit formulation loaded into a micro chamber is shown above.

# Micro Chamber Test Conditions

---

Parameter	Value
Sample Diameter, m	0.064
Volume, m <sup>3</sup>	5.15E-05
Air Change Rate, h <sup>-1</sup> (N)	58.2
Loading, m <sup>2</sup> /m <sup>3</sup> (L)	62.5
Specific Air Flow Rate (N/L)	0.931
Initial Temperature	23 °C



# Spraying Generic Closed-Cell SPF for Holding Time Study

---



Applicator is shown on left and spraying equipment shown on right.

# Holding Time Study Sample Preparation and Storage

---



Five replicate closed-cell SPF samples are shown in the spray booth.



Sample substrate consists of cardboard sheets wrapped with clean aluminum foil.




Samples stored in polyethylene terephthalate (PET) layered bags during holding time study

# Collecting Emissions from Generic SPF Specimens in Micro Chambers

Sample Description	Time, min	Sample Collection and Conditions
Starting Point	0	30 minute equilibration at 23 °C
TD Sample Point 1	30	Thermal Desorption (TD) tubes Sample for 20 minutes (Volume = 1L)
DNPH Sample Point 1	50	Aldehydes with DNPH tubes Sample for 1 hour (Volume = 3L)
TD Sample Point 2	110	Thermal Desorption (TD) tubes Sample for 20 minutes (Volume = 1L)
PP Filter for MDI	130	MDI with 13mm filter with PP Sample for 16 hours
TD Sample Point 3	1090	Thermal Desorption (TD) tubes Sample for 20 minutes (Volume = 1L)
Elevate Temperature	1110	Heat chamber to 40 °C and hold for 10 minutes
TD Sample Point 4	1120	Thermal Desorption (TD) tubes Sample for 10 minutes (Volume = 0.5L)
DNPH Sample Point 2	1130	Aldehydes with DNPH tubes Sample for 1 hour (Volume = 3L)



# Sampling from Micro Chamber



13mm Glass-Fiber Filter with  
1-(2-pyridyl)piperazine (PP) & diethyl phthalate



90mm Glass-Fiber Filters with PP



BIOS DryCal® Calibrator



Markes TD Tubes



Silica Gel with DNPH Tube



# Holding Time Observations

---

Generic closed-cell and open-cell SPF samples can be stored for up to 48 hours at 23 °C in sealed PET bags without causing a significant change in the emission factor values for target compounds known to be present in the formulation.

The emission factor for the amine catalyst in the low-pressure kit formulation did not appear to be stable in the stored samples. Further research may be necessary for this particular formulation.


# MDI Evaluations during Holding Time Study

---

The MDI was reported as non-detect for all samples collected during the holding time studies:

Component	$\mu\text{g}/\text{m}^3$
2,4' MDI	<2
4,4' MDI	<2
Polymeric MDI	<60

No MDI detected on walls



# SVOA Wall Effects Evaluation Spiking Study


---

There is a concern that semi-volatile organic compounds may adhere to the walls of the environmental test chambers, which could significantly bias the emission results.

The SVOA compounds of interest (MDI, selected amine catalysts and flame retardant) were spiked into micro chambers and small-scale stainless steel and PTFE lined chambers to evaluate recoveries.

# MDI Spike Study with Micro Chamber to Evaluate Wall Effects

---



- MDI (generic formulation) weighed into PTFE coated tray, then placed into micro chamber
  - Flow rate = 100 cc/min
  - Sample temperature and time = 65°C for 2 hours



- MDI emissions collected on 13mm PP filters
- Chamber wall and lid sampled for MDI using 90mm PP filter wipes

# MDI (Mixture) Spike Study Results

---

*Results reported as % MDI found on chamber walls:*  
(MDI on Wall) / (MDI on filter + MDI on wall) x 100

Chamber	2,4'-MDI	4,4'-MDI	p-MDI
1	4.6%	8.2%	ND
2	9.4%	37.4%	ND
3	9.7%	28.0%	ND
Blank	ND	ND	ND

ND = Not detected in chamber wall or PP filter.

# Heated Injector System for Spiking Small-Scale and Micro Chambers


---



The Supelco Adsorbent Tube Injector System employs the technique of flash vaporization to vaporize the sample into a continuous flow of an inert gas, which carries the sample to the test chamber or sorbent tube.

# Micro Chamber with 13mm PP Filter to Evaluate MDI Recovery


---






# Micro Chamber Spike with XAD-2 Tube

---




# Stainless Steel Chamber Spike

---



# PTFE Lined Acrylic Chamber Spike

---



# Spiking Study Observations

---

MDI spike recoveries were not consistent and a significant percentage of the spiked MDI adhered to the chamber walls, regardless of the material and size of the test chamber.

Amine catalyst recoveries were acceptable with the micro chamber (ranged from 81 to 99%).

The flame retardant was fully recovered in the micro chamber using air as the carrier gas (containing humidity) with extended sampling time.

# Summary of SPF Emissions Research

---

Developed thermal desorption GC/MS method for measuring potential SPF emissions

Established baseline conditions and sample preparation techniques for measuring SPF emissions with the micro-chamber

Conducted SPF insulation holding time and packaging study using generic SPF formulations; ASTM draft

Evaluated wall effects of semi-volatile compounds in small-scale and micro chambers

# Thank you!

---

## Questions

